


Thrombektomie Instrument - Dispositif de thrombectomie - Dispositivo per trombectomia - Instrumento para trombectomia
- Instrumento para trombectomia - инструмент для тромбэктомии

phenox GmbH
Lise-Meiner-Allee 31 - 44801 Bochum, Germany
Tel: +49 (0) 234 / 36919-0 - Fax: +49 (0) 234 / 36919-19
www.phenox.net

6 mm x 30 mm
Shaft Ø Working Length

ID 0.021" 3 mm
Microcatheter Min. Vessel Ø



A: Shaft Diameter [mm]	B: Shaft Length [mm]	C: Insertion Wire Length [mm]	D: Marker Distance [mm]	E: Working Length [mm]	Microcatheter Compatibility [Inch]	Min. Vessel Diameter [mm]
6	48	1800	1275	30	ID 0.021"	3

REF PRE-6-30 LOT 123456 2021-12 CONT 1 x pRESET STANDARD

pRESET STANDARD REF PRE-6-30 LOT 123456



(01)04260123780128(17)211231(10)123456

pRESET STANDARD phenox GmbH: www.phenox.net

REF PRE-6-30
LOT 123456 2021-12
(01)04260123780128(17)211231(10)123456

pRESET STANDARD phenox GmbH: www.phenox.net

REF PRE-6-30
LOT 123456 2021-12
(01)04260123780128(17)211231(10)123456

CE
0297

pRESET STANDARD phenox GmbH: www.phenox.net

REF PRE-6-30
LOT 123456 2021-12
(01)04260123780128(17)211231(10)123456

pRESET STANDARD phenox GmbH: www.phenox.net

REF PRE-6-30
LOT 123456 2021-12
(01)04260123780128(17)211231(10)123456



Phenox Inc.

Laser-free

NITI

Sterilized with Ethylene Oxide Gas

Consult instructions for use

Caution

Do not reuse

Keep dry

Keep away from sunlight

Do not use if the packaging is damaged

Do not sterilize

Non pyrogenic

LAB-B825-PRESET F - 201901



REF PRE-6-30

LOT 123456

2021-12