

Fox sv: Box- and pouch-label new (OUS)

Fox sv PTA Catheter (01)00000000000000(17)010131(10)000000(91)3000


REF LV_ITNR LOT *VOID LV_EXPD

∅ BALLOON 2.0 mm ↔ BALLOON 20 mm ↔ USAB CAT 90 cm SHEATH 4 Fr 1.33 mm

Fox sv PTA Catheter

REF LV_ITNR
 LOT *VOID
 LV_EXPD

CONTENTS 1 Fox sv OTW PTA Catheter



∅ BALLOON

2.0 mm

↔ BALLOON

20 mm

RBP

22 bar
2200 kPa

↔ USAB CAT

90 cm

SHEATH

4 Fr
1.33 mm

REC GW

0.018"
0.46 mm



STERILE R

NON PYROGENIC



15°C / 30°C



DIST

Abbott Vascular Devices Holland B.V.,
 Maastricht Airport, Bamfordweg 1,
 NL-6235 NS Uilestraten, The Netherlands
 Phone: +31 43 358 6751
 FAX: +31 43 358 6759



Abbott Laboratories Vascular Enterprises Ltd.
 Dublin, Beringen Branch, Anspottstrasse,
 CH-8222 Beringen, Switzerland,
 Phone: +41 52 674 8550
 FAX: +41 52 674 8540



© 2008 Abbott Laboratories

LA 009171001 005


DO NOT EXCEED!

RBP

**

NPR

Balloon Compliance			∅ 2.0
Pressure	Diameter		mm
	kPa	bar	
600	6	1.8	
700	7	1.9	
800	8	1.9	
900	9	1.9	
1000	10	2.0	
1100	11	2.0	
* 1200	12	2.0	
1300	13	2.0	
1400	14	2.0	
1500	15	2.1	
1600	16	2.1	
1700	17	2.1	
1800	18	2.1	
1900	19	2.1	
2000	20	2.1	
2100	21	2.1	
** 2200	22	2.2	
2300	23	2.2	
2400	24	2.2	
2500	25	2.2	
2600	26	2.2	
2700	27	2.3	
2800	28	2.3	
2900	29	-	
3000	30	-	




Fox sv PTA Catheter ∅ 2.0 mm ↔ 20 mm



Fox sv PTA Catheter ∅ 2.0 mm ↔ 20 mm

REF LV_ITNR LOT *VOID ↔ BALLOON 20 mm
 ALVE Ltd, Dublin, Beringen Branch, CH-8222 Beringen,
 Phone: +41 52 674 8550 FAX: +41 52 674 8540

Fox sv: Compliance chart new (OUS/US)



Fox sv PTA Catheter				
Balloon Compliance \varnothing 2.0 \pm .20				
	Pressure			Diameter
	[kPa]	[bar]	[atm]	[mm]
	600	6	6	1.8
	700	7	7	1.9
	800	8	8	1.9
	900	9	9	1.9
	1000	10	10	2.0
	1100	11	11	2.0
*	1200	12	12	2.0
	1300	13	13	2.0
	1400	14	14	2.0
	1500	15	15	2.1
	1600	16	16	2.1
	1700	17	17	2.1
	1800	18	18	2.1
	1900	19	19	2.1
	2000	20	20	2.1
	2100	21	21	2.1
**	2200	22	22	2.2
	2300	23	23	2.2
	2400	24	24	2.2
	2500	25	25	2.2
	2600	26	26	2.2
	2700	27	27	2.3
	2800	28	28	2.3
	2900	29	29	-
	3000	30	30	-

*Nominal Pressure **RBP Do not exceed
 ** NPR
 LOT *VOID

LA 009172001 003

DO NOT EXCEED!

RBP

**

NPR

Fox sv: Box new



Fox sv: Blister new

Embossed Abbottlogo on blister adjusted as follows:

