

DECLARATION OF CONFORMITY**Manufacturer:** Abbott Vascular**Address:** 3200 Lakeside Drive
Santa Clara, California 95054, USA**Additional
Manufacturing Sites:** 26531 Ynez Road
Temecula, California 92591, USA**Device Name:** **HI-TORQUE SUPRA CORE[®] 35 Guide Wire****Device Classification:** Class III**Classification Rationale:** The following Annex IX definition(s) apply to the HI-TORQUE SUPRA CORE[®] 35 Guide Wire for purposes of classification: Per Rule 6, Annex IX, all surgically invasive devices intended for transient use are in Class IIa unless they are: intended specifically to control, diagnose, monitor, or correct a defect of the heart or of the central circulatory system through direct contact with these parts of the body, in which case they are in Class III.**Authorized European
Representative:** Abbott Vascular International BVBA
Park Lane
Culliganlaan 2B
1831 Diegem, Belgium

Model Numbers: HI-TORQUE SUPRA CORE® 35 Guide Wire

Description	Part Number
HI-TORQUE SUPRA CORE 35, 145 CM	1002703
HI-TORQUE SUPRA CORE 35, 190 CM	1002703-01
HI-TORQUE SUPRA CORE 35, 300 CM	1002703-02

I, the undersigned, hereby declare that the medical devices specified above conform with the applicable *Essential Requirements* listed in Annex I and Annex II Part 4 of EC Council Directive 93/42/EEC.

This declaration is supported by an EC quality system and design examination approval (Annex II) under the supervision of the British Standards Institution, a notified body authorized by the United Kingdom Competent Authority. (EC Quality System Certificate Number CE 510108; EC Certificate Number CE 518028).

By: Susan Slane Date: 5/24/10
Susan Slane
Divisional VP, Quality, Compliance, and Analytical Chemistry
Abbott Vascular